

| Annexure Specification of Low sulfur Analyzer | | To Be Filled by vendor |
|--------------------------------------------------|--------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|------------------------|
| Test Method | Fully compliance with ASTM D-5453, UOP 987 | |
| Range | ASTM D5453 => 1.0 mg/kg to 8000 mg/kg total sulfur UOP 987 => 10 to 1500 ng/g (mass-ppb) total sulfur | |
| Operating temperature | Below 10- 30 °C | |
| Accessories & Requirements | All required accessories must be supplied | |
| Standard Accessories | All standard accessories must be included in the offer | |
| Consumables | A kit of basic consumables for two years trouble free operation must be included in the offer | |
| Calibration & Reference Standard | Certified & Reference standards specially designed and produced for on-site qualification of equipment during installation and must be as per method and comply with ISO 17034 | |
| Installation & Commissioning | Must be installed & commissioned by manufacturer's representative | |
| Training | A comprehensive On-site training from Trained manufacture representative to our staff is required | |
| Required Documents | Operation Manual | |
| | Service Manual (including all Electrical, Electronics, & Mechanical schematics) | |
| | Factory Calibration Certificate | |
| | Factory Test Certificate | |
| | Certificate of Origin | |

Handwritten signature: *Kat... as per...*